

Test report - PMFT 1000 - Oil penetration



Operator	Date	Time
MaB	21.12.2020	17:49:20

Additional information

future live AD GmbH Untere Vorstadt 12, 78532 Tuttlingen fensmed@gmx.de

Basic information

Operation parameter		Test conditions	
Script	Efficiency Test (oil)	Temperature	21.54 °C
Aerosol	Oil	rel. humidity	3.06 %
Main air	94.96 l/min	Mass concentration	17.54 mg/m³
Generator air	1.74 l/min		


Mask information

Mask ID	0145_20201221_174920
Mask los	TexTech Maske
Mask type	FFP 2

Result

Penetration	Measurement	Upper limit	Result
P_SARS-CoV-2 (0.145 µm)	8.344 %		
P_m	0.391 %		
P_Ph_45°	0.868 %	6.00 %	✓ pass

Distributions



Calibration valid until 25.08.2021