

Test report - PMFT 1000 - Breath resistance / Valve test



Operator	Date	Time
MaB	21.12.2020	17:44:27

Additional information

future live AD GmbH Untere Vorstadt 12, 78532 Tuttlingen fensmed@gmx.de

Basic information

Operation parameter	
Script	Breath Resistance
Aerosol	Pressurized air

Test conditions	
Temperature	21.54 °C
rel. humidity	3.52 %

Mask information

Mask ID	0145_20201221_174427
Mask los	TexTech Maske
Mask type	FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.05 l/min	52.70 Pa	70.00 Pa	✔ pass
Inhalation 2	95.40 l/min	170.48 Pa	240.00 Pa	✔ pass
Exhalation	160.55 l/min	290.46 Pa	300.00 Pa	✔ pass



Calibration valid until 25.08.2021